THE INSTITUTE OF PETROLEUM PETROLEUM MEASUREMENT MANUAL

PART X Meter Proving

SECTION 7

CALIBRATION OF REFERENCE METERS USED FOR GANTRY METER PROVING

SUB-SECTION 7.1

SMALL VOLUME PROVER METHOD



INSTITUTE OF PETROLEUM PETROLEUM MEASUREMENT MANUAL

PART X Meter Proving

SECTION 7

CALIBRATION OF REFERENCE METERS USED FOR GANTRY METER PROVING

SUB-SECTION 7.1

SMALL VOLUME PROVER METHOD

August 1996

Published by
The Institute of Petroleum, London
A charitable company limited by guarantee

Copyright $\mathbb O$ 1996 by The Institute of Petroleum, London: A charitable company limited by guarantee. Registered No. 135273, England

All rights reserved

No part of this book may be reproduced by any means, or transmitted, or translated into a machine language without the written permission of the publisher.

ISBN 0 85293 171 9

Printed by The Institute of Petroleum



Petroleum Measurement Manual, Part X: Meter Proving Section 7: Calibration of Reference Meters Used for Gantry Proving.

Addendum

This section of the Petroleum Measurement Manual was published as two sub-sections:

- 7.1: Small Volume Prover Method (1996)
- 7.2: Volumetric and Gravimetric Tank Method (1999)

Both sub-sections refer to the fact that a reference meter should be calibrated on each specific type of product that will be measured by the gantry meters to be proved by the reference meter. Despite this, it is acknowledged that reference meters have sometimes been calibrated on products similar to those on which they will subsequently be used, and a fixed correction applied to the meter factor to account for the change in performance of the meter.

Recent research, sponsored by the Institute of Petroleum, has demonstrated the magnitude of the errors that may be generated by following this approach.

Reference PD meters are, on occasions, calibrated on gas oil and then used on ultra low sulphur diesel (ULSD), with a meter factor correction of 0,1% applied to obtain the meter's performance on ULSD. Tests on six reference PD meters showed significant variations in performance between individual meters, with the result that the use of a fixed correction of 0,1% would have produced metering errors of up to 0,07% at 2250 l/min (the maximum flow rate), with larger errors at lower flow rates.

Similar performance variations between individual meters were noted with motor spirit. Such variations will lead to significant errors when a reference PD meter is calibrated on a motor spirit substitute and a fixed meter factor correction applied to establish its performance on motor spirit.

The recommendation in the original documents, that a reference meter should be calibrated on each specific type of product on which it will subsequently be used, is therefore reinforced by the results of recent IP research.



CONTENTS

			Page	
Forewo	ord		. vii	
Acknow	wledgen	nents	viii	
1	Introd	luction and Scope	1	
2	Glossa	ary of Terms	2	
3	Safety	Precautions	5	
4	The R	eference Meter	6	
	4.1	Selection of Reference Meters		
	4.2	The Pulse Generator		
	4.3	General Considerations		
5	Measurement Standards			
	5.1	Small Volume Provers		
	5.2	Gravimetric Standards		
	5.3	Primary Volumetric Measures		
6	Ancillary Equipment and Facilities			
	6.1	Product Recirculation Facilities		
	6.2	Temperature Measurement		
	6.3	Pressure Measurement		
7	Trace	ability to National Standards	. 12	
8	Calibration of the Small Volume Prover			
	8.1	General Considerations		
	8.2	Gravimetric Water-Draw Method		
	8.3	Volumetric Water-Draw Method		
	8.4	Correction Factors		
	8.5	Small Volume Prover Calibration Certificate	. 22	

Contents Cont....

		age
9.1	General Considerations	23
9.2	Frequency of Calibration	23
9.3	Installation - Mechanical	24
9.4	Installation - Electrical	24
9.5		
9.6		
9.7	Achieving Temperature Stability	25
9.8	Control System Checks	25
9.9		
9.10	Flowrate Selection and the Minimum Number of Flowrates Required	25
9.11		
9.12	Temperature and Pressure Measurements During a Proving Run	26
9.13	Data Acquisition and Record Keeping	26
9.14		
9.15		
9.16	Reference Meter Calibration Certificate	28
ces		
Corre	ection Factors	29
A.1		
A.2		
A.3	Factors for Liquid Compressibility	30
A.4	Factor for Thermal Expansion of Metal	30
A.5	Factor for Expansion of SVP Due to Pressure	31
Exam	ple of a Reference Meter Calibration Certificate/Correction Table	32
	9.1 9.2 9.3 9.4 9.5 9.6 9.7 9.8 9.9 9.10 9.11 9.12 9.13 9.14 9.15 9.16 Corre A.1 A.2 A.3 A.4 A.5	Calibration of the Reference Meter 9.1 General Considerations 9.2 Frequency of Calibration 9.3 Installation - Mechanical 9.4 Installation - Electrical 9.5 Priming and Venting 9.6 Seal Integrity Checks 9.7 Achieving Temperature Stability 9.8 Control System Checks 9.9 Minimum Number of SVP Passes and Runs Required 9.10 Flowrate Selection and the Minimum Number of Flowrates Required 9.11 Flowrate Constancy Criteria 9.12 Temperature and Pressure Measurements During a Proving Run 9.13 Data Acquisition and Record Keeping 9.14 Reference Meter Calibration Procedures 9.15 Calculation and Application of Correction Factors 9.16 Reference Meter Calibration Certificate Correction Factors A.1 Nomenclature A.2 Factors for Liquid Thermal Expansion A.3 Factors for Liquid Compressibility A.4 Factor for Thermal Expansion of Metal

FOREWORD

Measurement accuracy is essential for the sale, purchase and handling of petroleum products. It reduces the likelihood of disputes between buyer and seller and facilitates the control of losses. Accurate measurement demands the use of calibration equipment with accuracy traceable to national or international standards and the adoption of standard procedures.

The Petroleum Measurement Committee of the Institute of Petroleum is responsible for the production and maintenance of standards and guides covering the various aspects of static and dynamic measurement of petroleum. These are issued as separate Parts and Sections of the Institute's *Petroleum Measurement Manual*, which was first published in 1952.

Membership of the IP working panels is made up of experts from the oil industry, equipment manufacturers, cargo inspectors and government authorities. Liaison is maintained with parallel working groups of the Committee on Petroleum Measurement of the American Petroleum Institute, and is extended as necessary to embrace other organizations concerned with quantitative measurement in other countries and in other industries.

Users are invited to send comments, suggestions, or details of experience with this issue to:

The Technical Manager, Petroleum Measurement Institute of Petroleum 61 New Cavendish Street London W1M 8AR United Kingdom

The Petroleum Measurement Manual is widely used by the petroleum industry and has received recognition in many countries by consumers and the authorities. In order to promote their wide adoption internationally, it is the policy to submit selected standards through the British Standards Institute to Technical Committee TC 28 - Petroleum Products and Lubricants - of the International Organization for Standardization (ISO/TC 28) as potential International Standards.

A full list of the Parts and Sections of the *Petroleum Measurement Manual* (PMM) is available on request from the Institute of Petroleum.

Note

The IP Petroleum Measurement Manual is recommended for general adoption but shall be read and interpreted in conjunction with weights and measures, safety and other regulations in force at the particular location where it is to be applied. Such regulatory requirements shall have precedence over the corresponding clauses in the Manual except where the requirements of the Manual are more rigorous, when its use is recommended. The Institute disclaims responsibility for any personal injury or loss or damage to property, howsoever caused, arising from the use or abuse of any Part or Section of the Manual.

ACKNOWLEDGEMENTS

This Part of the Petroleum Measurement Manual was prepared by a Working Group drawn from the Commercial Metering Committee, PM-D-4, which is part of the Institute's Petroleum Measurement Committee structure.

Members of the Working Group associated with production of the standard were:-

A M Babb

Avery Hardoll Fluid Management

P A Barlow

SGS Redwood Ltd

A Butcher

FMA Ltd

D Helks

BP Oil UK Ltd

J D Snook

Consultant

A J Thorogood Consultant

J M Wood

Institute of Petroleum (Editorial)

Company affiliations are those that applied at the time the document was being drafted.

1

INTRODUCTION AND SCOPE

Recommended procedures for the proving of gantry meters are provided in Part X of the IP Petroleum Measurement Manual, Sections 2 and 5, which describe the use of reference meters and volumetric proving tanks as traceable reference measurement standards.

While the earlier sections of Part X of this Manual provide general guidance on the various methods of reference meter calibration, no detailed procedures for calibration have hitherto been available. Methods for the calibration of reference meters include the use of gravimetric facilities within calibration laboratories, or the use of volumetric proving tanks or pipe provers under either laboratory or field conditions.

This sub-section of the Petroleum Measurement Manual has been prepared to provide guidance for the calibration, using small volume provers (SVPs), of reference meters specifically used for proving loading gantry meters. A principal requirement for this application is that the proving be carried out, wherever practicable, using the actual grades of product metered by the gantry meters, in order to minimise measurement uncertainty associated with the use of different viscosity test liquids.

This procedure for calibration of the reference meter using SVPs may therefore differ from that described in other Parts of the Petroleum Measurement Manual dealing with other applications of small volume provers and their calibration.

Small volume provers provide a closed system method of meter calibration that may be used in the field. They offer the advantage that comparatively small volumes of actual metered products are used in the operation.

This document describes the selection and calibration of equipment and the operational procedures and calculation routines that are recommended when using SVPs. Procedures for the calibration of reference meters with volumetric proving tanks will be detailed in a future sub-section 7.2 of Part X of the Petroleum Measurement Manual.

It should be noted that the method described herein is intended only for use with refined products that are loaded into non-pressurised road or rail vehicles via gantry meters. It is not intended for use with products such as live crude oil or liquefied petroleum gases.

The use of the verb 'shall' in this document indicates an action that is required to be observed in order to achieve the high standards of measurement accuracy and traceability achievable by this procedure.